

The most reliable LC-EC applications for Drugs & Pharmaceuticals analysis

**Antipsychotic drugs**

Clozapine  
Olanzapine  
Risperidone

**PET imaging tracer**

Fluorodeoxyglucose (FDG)  
FDG impurities

**Pharmaceuticals, API**

Acetaminophen  
Artemether  
Artemisinin  
Dihydro- artemisinin  
Betadex sulfobutyl ether sodium  
Etoposide  
Epinephrine  
Heparin  
mesna BNP7787  
8-OH-DPAT  
Vincristine

**Sulfides**

Glutathione  
Aminothiols  
Disulfides

**Aminoglycoside drugs**

Amikacin  
Framycetin sulphate  
Gentamicin sulphate  
Kanamycin  
Netilmycin  
Neomycin sulfate  
Spectinomycin  
Lincomycin  
Tobramycin

## Spectinomycin and Lincomycin

- **FlexCell with exchangeable gold electrode**
- **Analysis of main substituent and impurities**
- **Reproducible & robust**

### Summary

In this note a method is described for the simultaneous analysis of spectinomycin and lincomycin using the ALEXYS analyzer based on the Antibiotics LC base system with post-column addition kit. The method is based on reversed phase chromatography with a step gradient. Detection is accomplished using post-column addition of sodium hydroxide in combination with pulsed amperometric detection (PAD).

## Introduction

Lincomycin and spectinomycin (Fig. 1) are aminoglycoside antibiotics that are mainly used for veterinary purposes. They are often added as a mixture to the drinking water of poultry to prevent respiratory infections. The simultaneous analysis of both components in formulations is complicated by the large difference in chromatographic retention behavior [1].

In this note an LC-ECD method is presented for the simultaneous analysis of spectinomycin and lincomycin using the ALEXYS analyzer.


## Method

### Separation

The chromatography is based on a step gradient elution using two mobile phases that differ in ionic strength and THF concentration (Table 1).

The composition of mobile phase A is chosen so that spectinomycin and the early eluting impurities are separated within 15 minutes (Fig. 2). Both components contain amine groups (which are protonated at pH 3) and their retention on a C18 column is facilitated by ion pairing with octane sulfonate.

To speed up the elution of lincomycin, mobile phase B is applied after 12 minutes. The retention time of lincomycin is affected by the change in THF (Fig. 3) but also by ionic strength (which affects the interaction with the ion pairing agent). We found that a THF concentration higher than 2% should not be used, as this results in mobile phase precipitation (it turns a milky white color), which will clog the system. It is also important to use stabilized THF to assure low cell currents. After the elution of lincomycin, the system is allowed to stabilize for 26 minutes in mobile phase A.




**Figure 2:** Analysis of 100 mg/L spectinomycin (1) and lincomycin (2) dissolved in mobile phase A (red trace) in an overlay with the analysis of a blank (blue trace). LC-ECD conditions as in Table 1. The bottom bar indicates the timing of

**Table 1**


### LC-ECD conditions (EP)

HPLC*	ALEXYS Antibiotics base system - Isocratic + Post-column kit EP
Column	Thermo Scientific™ Betabasic™ 18 HPLC Column 150x3.0mm, 3 μm
Mobile phase (MP)	<b>A:</b> 25 mM citric acid, 25 mM phosphoric acid, pH 3.0 (set with NaOH solution), 25 mg/L octanesulfonic acid sodium salt, 2 mL/L stabilized THF <b>B:</b> 100 mM citric acid, 100 mM phosphoric acid, pH 3.0 (set with NaOH solution), 25 mg/L octanesulfonic acid sodium salt, 20 mL/L stabilized THF
MP gradient	0 - 12 min: A; 12 - 34 min: B; 34 - 60 min: A
Flow rate	0.4 mL/mL
Post-column addition	1 M NaOH at 0.2 mL/min
Mixing coil volume	80 μL
Temperature	35 °C for separation, mixing and detection
V <sub>injection</sub>	20 μL
Flow cell	FlexCell™ with Au and HyREF, 50 μm spacer
Potential waveform (3-step)	E1, E2, E3: +0.12, +0.7, -0.7 V t1, t2, t3, ts: 0.4, 0.2, 0.35 s, 80 ms
Range	50 μA/V
ADF	0.5 Hz
I-cell	About 7 μA


\*) Note that the presented data are obtained with an older version of the ALEXYS system than shown in figure 9.



**Figure 1:** Structural formulas of lincomycin (l) and spectinomycin (r)



**Figure 3:** Effect of THF concentration in mobile phase on the retention time of spectinomycin (blue) and lincomycin (red) when running an isocratic separation.



**Figure 4:** Retention time of lincomycin after prolonged equilibration with mobile phase B before t=0. The first 3 runs show a strong shift in retention time due to equilibration of the LC system (red). If a sequence is preceded by running the complete step-gradient 3 x, then the initial stability is greatly improved (blue).

## Chromatographic equilibration procedure


Before starting a sequence of analyses, the ion pairing chromatographic conditions have to be equilibrated, otherwise retention will shift during the first 3 analyses (Fig. 4). The stabilization takes about 3 hours, during which the gradient should be run 3 times. A comparable pattern was observed for spectinomycin (not shown).

## Detection

Electrochemical detection of lincomycin and spectinomycin requires a high pH (post-column NaOH addition with 2<sup>nd</sup> pump) and Pulsed Amperometric Detection (PAD) on a gold electrode. A 3-step pulse was applied (Table 1).

## Flow cell maintenance

As the gold working electrode is slowly consumed during 3-step pulsed amperometric detection, the cell volume increases, which leads to lower signals over time. To restore the signal and cell volume, the gold electrode should regularly be flattened and polished to the original flat surface [5]. The FlexCell has a removable electrode and is easy to service. After polishing the gold electrode, the system needs 10 h of stabilization time while running the gradient program (Fig. 5) to reach good prolonged signal stability with an intraday RSD of 2% or better (see results section). We recommend to clean the gold electrode at least every 2 weeks when in continuous use.



**Figure 5:** Relative peak area of lincomycin after having polished the gold working electrode. Blue squares represent a series of continuous analyses, and red circles represent a series of analyses where blank gradient traces were run for 10 hours between the first and subsequent injection.

## Results

A typical chromatogram as obtained with the method is shown in Fig. 2. A lag time of about 4 minutes can be observed when comparing the changes in the baseline and the step gradient pattern. Spectinomycin and lincomycin are well separated from blank and baseline disturbances. The early eluting impurities from spectinomycin are separated within the first 15 minutes (Fig. 6).




Figure 6: Detail of Fig. 2 showing the first 15 minutes where spectinomycin impurities elute.




Figure 7: Calibration plot with linear regression lines for spectinomycin and lincomycin.

## Linearity

In the concentration range of 10-50 mg/L the correlation coefficient with peak area is 0.999 or better for spectinomycin as well as lincomycin (Fig. 7).

## Intra-day reproducibility

A representative overlay of 6 consecutively measured chromatograms is given in Fig. 8 with RSD values given in Table 2. Reproducibility (n=6) of 2% RSD or better was observed for peak area of spectinomycin and lincomycin on all days. This RSD value is better than the system requirements for the comparable isocratic analysis of spectinomycin, according to the European Pharmacopoeia [3], which is 3% RSD (n=6) or better. Reproducibility of 0.2% or better was observed for retention times.

Table 2

Reproducibility (chromatograms in Fig. 8); n=6

Component	RSD (%) retention time	RSD (%) height	RSD (%) area
Spectinomycin	0.18	2.0	2.0
Lincomycin	0.08	1.3	2.0




Figure 8: Overlay of 6 chromatograms of 100 mg/L spectinomycin and lincomycin in mobile phase A.



## Inter-day reproducibility

The inter-day repeatability was measured in relation to the gold electrode maintenance procedure (see method section). Average peak response data from 4 different days are presented in Table 3. For each day, different events preceded the analysis of data: on day 1 and 12 the electrode was polished before analyses, on day 11 the analysis were done after having continuously run the gradients (stand-by mode), and on day 10 the system had been left to equilibrate for 3 hours (running gradients) after a system shut down of a day. Variation in peak area or peak height (intra-day repeatability) was better than 6% RSD.

Table 3

Interday reproducibility.  
Average response (n=6) on different days, based on analyses of 100 mg/L spectinomycin and lincomycin in mobile phase A

Day	Events before analysis	Peak height ( $\mu A$ )		Peak area ( $\mu A \cdot sec$ )	
		Spec.	Linco.	Spec.	Linco.
1	Polish + 10 h equilibration	8.4	6.0	298	255
10	Shut down + 3 h equilibration	9.6	6.7	290	264
11	Continuous analyses	8.6	6.7	263	272
12	Polish + 12 h equilibration	9.1	6.4	293	275
Intra-day average		9	6	286	267
RSD (%)		5.9	5.7	5.5	3.5

## References

1. J. Szúnyog, E. Adams, K. Liekens, E. Roets, J. Hoogmartens, *Journal of Pharmaceutical and Biomedical Analysis* 29:213-220 (2002)
2. W. R. LaCourse, *Pulsed Electrochemical Detection in High-Performance Liquid Chromatography*, Wiley, New York, 1997
3. "Spectinomycin Dihydrochloride Pentahydrate", *European Pharmacopoeia*, 6.0, (2008) 2947-2949
4. V. P. Hanko, W. R. LaCourse, C. O. Dasenbrock, J. S. Rohrer, *Drug Development Research* 53:268-280 (2001).
5. Antec Scientific, Flattening & Polishing kit for metal WE: User Guide, part number 250.7010

## Conclusion

The ALEXYS analyzer for spectinomycin and Lincomycin is a robust and reliable solution for the routine analysis of spectinomycin, lincomycin and its impurities in pharmaceutical formulations.



## Ordering information

<b>Recommended ALEXYS analyzer + parts</b>	
180.0058W	ALEXYS Antibiotics base system - Isocratic
180.0605EP	Post Column Kit EP
102.4325EP	Flexcell Au HyREF with stainless steel AUX
250.1045	Flattening/polishing kit for metal WE
184.0209	Glass bottle assembly, 1L, Helium

**Figure 9.** The ALEXYS analyzer for Spectinomycin and Lincomycin, consisting of the ALEXYS Antibiotics base system - isocratic, post-column addition kit (NaOH) and FlexCell with gold working electrode, HyREF and stainless steel auxiliary electrode. The ALEXYS analyzer is fully controlled by DataApex™ Clarity™ software.

**For research purpose only.** The information shown in this communication is solely to demonstrate the applicability of the ALEXYS system and DECADE Elite detector. The actual performance may be affected by factors beyond Antec's control. Specifications mentioned in this application note are subject to change without further notice.

DECADE Elite, ALEXYS, SenCell, FlexCell and HyREF are trademarks of Antec Scientific. Clarity™ and DataApex™ are trademarks of DataApex Ltd. Chromeleon™ is a trademark of Thermo Fisher Scientific. OpenLAB™ and Chemstation™ are trademarks of Agilent Technologies, Inc. All other trademarks are the property of their respective owners.

**Antec Scientific (USA)**  
 info@AntecScientific.com  
 www.AntecScientific.com  
 T 888 572 0012

**Antec Scientific (worldwide)**  
 info@AntecScientific.com  
 www.AntecScientific.com  
 T +31 71 5813333

