



The finest HPAEC-PAD
Applications for
Carbohydrate analysis

Food and Beverage

Mono- and disaccharides
Sugars in meat & fish
Carbohydrates in food
according to AOAC
Carbohydrates in instant
coffee
Carbohydrates in honey

Prebiotics Food additives

Fructans in infant formula
TGOS in food products
Profiling of FOS

Lactose Free Products

Lactose in dairy & meat
Lactose in lactose-free
products

Artificial Sweeteners

Sugar alcohols

Glycans & Glycoproteins

Monosaccharides in
glycoproteins
Analysis of N-glycans

Analysis of Maltodextrin in Syrups

- **ALEXYS™ Carbohydrate Analyzer**
- **SweetSep™ AEX200 anion-exchange column**
- **High-resolution separation for maltodextrin profiling**
- **Glucose syrup and high maltose corn syrup samples**

Summary

Maltodextrin is a carbohydrate mixtures, consisting of mono-, di-, and oligosaccharides. Maltodextrin is produced from starch using either chemical or enzymatical hydrolysis [1, 2]. Due to a large number of oligosaccharides, maltodextrin is often used as a nutritious additive in many food products, such as in dried instant food formulations, confectionary, cereals, snacks, syrups, and beverages [3]. However, maltodextrin has a high glycemic index, leading to concerns about increased blood glucose levels particularly for individuals with diabetes [4, 5]. Therefore, there is a demand for an analytical method that provides insights into maltodextrin profile and allows for sensitive quantification.

In this application note a method is presented for the analysis of maltodextrin using the ALEXYS™ Carbohydrate Analyzer. High-performance anion-exchange chromatography in combination with pulsed amperometric detection (HPAEC-PAD) is the method of choice for selective and sensitive analysis of maltodextrin. High-resolution separation of maltodextrin up to DP₄₀ were achieved using a strong anion-exchange column SweetSep™ AEX200. Pulsed amperometric detection enables sensitive detection with low detection limit. Two syrup samples were analyzed to demonstrate the applicability of the method.



Introduction

Maltodextrins are mixtures of carbohydrates obtained by partial hydrolysis of starch. As a product of partial hydrolysis of starch, maltodextrin consists of mainly oligosaccharides from a degree of polymerization (DP) of 3 up to 40, and a low amount of monosaccharides and disaccharides [1, 2]. The composition of maltodextrin is attributed to its low sweetness and nutritious value, which makes maltodextrin attractive to be used as an additive to food products [2, 6].

Despite the GRAS (Generally Recognized As Safe) status [1], maltodextrin is a substance with a glycemic index higher than table sugar [4, 5]. Consumption of food products modified with maltodextrin can therefore increase blood glucose, which is not preferred for people with diabetes. The sudden increase in blood glucose level is closely correlated to the monosaccharides and disaccharides amount in maltodextrin. Therefore, there is a need for an analytical method that can provide insights into the maltodextrin profile and sensitive quantification of maltodextrin.

A number of analytical methods can be used for the analysis of maltodextrin, such as using titration, or thin-layer chromatography [7–9]. However, these methods only produce qualitative results. In order to have reliable quantitative methods for the analysis of maltodextrin, high-performance

liquid chromatography with refractive index (RI) detection [9], or evaporative light scattering detection (ELSD) has been developed [10]. However, these methods are lacking of sufficient selectivity and sensitivity.

High-performance anion exchange chromatography in combination with pulsed amperometric detection (HPAEC-PAD) is the preferred method for maltodextrin analysis [11–13]. It combines superior selectivity with sensitive detection. The hydroxyl groups on carbohydrates can be oxidized, therefore, carbohydrates can be detected without derivatization using pulsed amperometric detection. In this application note, a method is presented based on a new type of anion-exchange column, the Antec Scientific SweetSep™ AEX200 with highly monodisperse 5 µm particles for the selective and sensitive analysis of maltodextrin.

Method

The method for the analysis of maltodextrin was evaluated using the ALEXYS™ Carbohydrate Analyzer (Figure 1). This dedicated HPAEC-PAD system consists of a P6.1L LPG quaternary pump, an AS6.1L autosampler, a CT2.1 column thermostat, an ET210 eluent tray for nitrogen blanketing, and the DECADE Elite electrochemical detector. Take into account



Figure 1: ALEXYS Carbohydrate Analyzer consisting of the ET210 eluent tray, P6.1L LPG quaternary pump, AS6.1L autosampler, CT2.1 column thermostat, and DECADE Elite electrochemical detector in combination with the SenCell™ 2 mm Au WE and HyREF (Pd/H₂) REF.

Table 1

| Conditions | |
|-----------------------------|--|
| HPLC system | ALEXYS™ Carbohydrate Analyzer |
| Detector | DECADE Elite electrochemical detector |
| Columns | SweetSep™ AEX200, 4 × 200 mm column, 5 µm SweetSep™ BIT, 4 × 50 mm borate ion trap All columns: Antec Scientific |
| Filter | Pre-column filter PEEK, 0.5 µm |
| Mobile phase | A: 100 mM NaOH B: 100 mM NaOH, 500 mM NaOAc Eluents prepared & blanketed with Nitrogen 5.0 |
| Flow rate | 0.7 mL/min |
| System backpressure | About 195 bar at the start of run, about 215 bar at the end of the NaOAc gradient (t=35 min) |
| Injection | 10 µL |
| Temperature | 30°C for separation, 35°C for detection |
| Flow cell | SenCell with 2 mm Au WE, stainless steel AE, and HyREF RE, AST 2 |
| Potential waveform (4-step) | E1, E2, E3, E4: +0.1, -2.0, +0.6, -0.1 V ts, t1, t2, t3, t4: 0.2, 0.4, 0.02, 0.01, 0.07 s |
| I cell | About 0.3-0.4 µA |
| ADF | 0.5 Hz |
| Range | 5 µA/V; 500 nA/V for LOD determination |



that the selection of a third-party quaternary HPLC systems may influence the separation performance, and may require some small adjustment of the conditions to achieve the results outlined in this application note. For sensitive detection of the carbohydrates, the SenCell™ with Au working electrode and HyREF (Pd/H₂) reference electrode was selected.

Separation

Carbohydrates are weak acids with pKa values ranging between 12 and 14. At high pH, carbohydrates will be either completely or partially ionized depending on their pKa value. Therefore, under alkaline conditions (pH > 12) carbohydrates can be separated using HPAEC. Only polymeric anion-exchange columns are suitable for carbohydrate separation at such alkaline conditions. The retention time of carbohydrates is inversely correlated with the pKa value and increases significantly with molecular weight. Maltodextrin is a breakdown product of starch consisting of unbranched linear glucose oligomer and polymer chains. Consequently, separation of such maltodextrin samples can be done based on their degree of polymerization. A high-resolution anion-exchange column with a 5 µm particle size, SweetSep™ AEX200, was chosen for the separation of carbohydrates in maltodextrin. The morphology and exchange capacity of this new AEX200 resin were specifically optimized to enable the analysis of a wide variety of carbohydrate samples, ranging from monosaccharides present in food, plants, and glycoproteins, up to polysaccharides such as in maltodextrin. The use of a pre-column filter is advised when using samples which might contain particulate matter.

The separation was performed at 30°C using the gradient profile shown in Table 2. The starting condition of the run is 100 mM NaOH with 40 mM sodium acetate. All carbohydrates were eluted in 30 minutes using a linear gradient from 40 mM up to 450 mM sodium acetate in 100 mM NaOH. The column

was subsequently cleaned for 5 minutes using 450 mM sodium acetate in 100 mM NaOH, followed by equilibration to the starting conditions for 15 minutes. The addition of a wash and an equilibration step resulted in a total run time of 50 minutes.

Detection

For the pulsed amperometric detection, a Sencell™ electrochemical flow cell is used [14]. This flow cell has a confined wall-jet design and consists of a Au working electrode (WE), a HyREF (Pd/H₂) reference electrode (RE), and a stainless steel auxiliary electrode (AE). A 4-step potential waveform was applied as described in Table 1. The detection temperature was set to 35°C. The cell current was typically about 0.3-0.4 µA under the specified conditions. This particular 4-step waveform has the following benefits (1): a consistent long-term peak area response and (2) minimal electrode wear [15], resulting in less flow cell maintenance and system down time.

Preparation of standards and samples

Standards: 50000 ppm stock standards of the individual sugars, glucose (DP₁), maltose (DP₂), and maltotriose (DP₃), and maltodextrin (DE* 4-7 and DE 10-15) were prepared in 95/5 (v/v%) water/acetonitrile. Acetonitrile was added to suppress bacterial and fungal growth and to prevent fast degradation. The stock standards under these conditions are stable for more than a month in the fridge at 4°C. A working standard mix of DP₁, DP₂, and DP₃, in the concentration ranges from 0.1 ppm to 10 ppm was prepared by serial dilution of the stock standards with DI water. The working standard of the maltodextrin (DE 4-7 and DE 10-15) was prepared by serial dilution of the stock standards with DI water, with a final concentration of 500 ppm. The concentration of the maltodextrin working standard is significantly larger than the DP₁—DP₃ mix due to the smaller response factor of higher DPs present in the maltodextrin compared to the DP₁—DP₃.

Samples: two types of syrup were obtained from a food ingredient company. One of the samples is a low sugar glucose syrup (GS), while the other sample is a high maltose corn syrup (HMCS). The stock solutions of the samples were prepared by weighing an appropriate amount of the syrups (GS = 208.1 mg, HMCS = 210.6 mg) and dissolving the syrups in 95/5 (v/v%) water/acetonitrile with a final concentration of 50000 ppm. Samples with a final concentration of 50 ppm were prepared by serial dilution of the stock solutions with DI water (dilution factor = 1000x).

Table 2

Gradient program

| Time (min) | Mobile phase | Description |
|------------|----------------------------|------------------------------------|
| 0 | 100 mM NaOH + 40 mM NaOAc | Elution & detection |
| 30 | 100 mM NaOH + 450 mM NaOAc | |
| 30-35 | 100 mM NaOH + 450 mM NaOAc | Column clean-up and regeneration |
| 35-50 | 100 mM NaOH + 40 mM NaOAc | Equilibration, starting conditions |

*) DE = Dextrose Equivalent is a measure of the amount of reducing sugars present in a sugar product, expressed as a percentage on a dry basis relative to dextrose. The dextrose equivalent gives an indication of the average degree of polymerisation (DP) for starch sugars.




Figure 2: Overlay chromatograms of the 10 μ L injections of the DP₁–DP₃ mix (black), maltodextrin DE4–7 (red), and maltodextrin DE 10–15 (blue).

Results

An overlay chromatogram of the 10 μ L injections of the DP₁–DP₃ mix, maltodextrin DE 4-7, and maltodextrin DE 10-15 is shown in Figure 2. DP₁–DP₃ are baseline separated (resolution > 1.5) and eluted within 5 minutes under this condition. It should be noted that the response factor is inversely proportional to the molecular weight, i.e. the higher the DP, the smaller the response factor.

Separations of the oligo- and polysaccharides in maltodextrin standards are excellent. In maltodextrin DE 4-7, a high-resolution separation of each DP up to DP₄₀ was observed within 25 minutes. In maltodextrin DE 10-15, the separation was achieved up to DP₃₀. Maltodextrin DE 4-7 theoretically consists of a larger proportion of the larger DPs (DP₁₄–DP₂₅) compared to Maltodextrin DE 10-15 (DP₆–DP₁₀). Therefore, it is obvious that the larger DPs in maltodextrin DE 4-7 have significantly higher responses compared to the maltodextrin DE 10-15.

Linearity

The linearity of the method was investigated for DP₁–DP₃. For higher DPs no linearity was measured due to the lack of commercial standards. The calibration curves for DP₁–DP₃ in the concentration range of 0.1–10 ppm are shown in figure 3. In this concentration range, the linearity is excellent, with




Figure 3: Calibration curves of DP₁, DP₂, and DP₃ in the concentration range of 0.1–10 ppm.

correlation coefficients larger than 0.9999 for DP₁, DP₂, and DP₃.

Repeatability

The repeatability of the method was evaluated by repetitive injections of a 0.1, 1 and 10 ppm DP₁ - DP₃ standard mix in DI water. The relative standard deviations (RSDs) for retention time and peak area for the three concentrations of the DP₁–DP₃ standards are listed in Table 3. Good repeatability of the method is evident from the low RSD values obtained. RSDs for retention time were < 0.25%.

Table 3

| Repeatability of 10 μ L injections of a 10 ppm and 1 ppm of DP ₁ –DP ₃ mix in water (n=10) | | | | | | |
|--|--------------------|------|-------------------|------|----------------------|------|
| Compound | RSDs (%) 10 ppm | | RSDs (%) 1 ppm | | RSDs (%)* 0.1 ppm | |
| | t_R | Area | t_R | Area | t_R | Area |
| DP ₁ | 0.18 | 0.13 | 0.14 | 1.41 | 0.18 | 2.27 |
| DP ₂ | 0.21 | 0.35 | 0.19 | 1.08 | 0.19 | 4.63 |
| DP ₃ | 0.13 | 0.36 | 0.09 | 2.23 | 0.15 | 2.79 |

*) n=5 for the repeatability test with 0.1 ppm standard.

For peak area, the RSDs were < 0.40% for all components in the 10 ppm standard and < 2.5% for all components in the 1 ppm standard. The RSDs of all components in the 0.1 ppm standard were < 5% because this concentration is close to the limit of detection (LOD). Nevertheless, the presented data demonstrate that a repeatable analysis of DP₁–DP₃ can be achieved.



LOD & LOQ

The limit of detection (LOD) and limit of quantification (LOQ) for DP₁—DP₃ in ppm level are shown in Table 4. The LODs were calculated as the analyte response corresponding to 3× the ASTM noise (average peak-to-peak baseline noise of 10 segments of 0.5 min). The noise was calculated based on a 5-minute section of the baseline. LOQs were calculated as the analyte response corresponding to 10× the ASTM noise. The average responses of 5 replicate injections obtained with a 0.1 ppm standard in the 500 nA/V range were used to calculate the LODs and LOQs for DP₁, DP₂, and DP₃. The sensitivity of the method is evident from Table 4. Higher DP maltodextrins have lower response factors. Therefore, DP₃ has the highest concentration detection limit (0.055 ppm) followed by DP₂ (0.040 ppm) and DP₁ (0.018 ppm), respectively.

Table 4

| Limit of Detections (LODs) and Limit of Quantifications (LOQs) | | |
|--|------------|------------|
| Compound | LODs (ppm) | LOQs (ppm) |
| DP ₁ | 0.018 | 0.058 |
| DP ₂ | 0.040 | 0.132 |
| DP ₃ | 0.055 | 0.184 |

Sample analysis

Two syrup samples labelled low sugar glucose syrup (GS) and high-maltose corn syrup (HMCS) were analyzed using the presented method. The analyses were performed to obtain the maltodextrin profile of the syrups and to quantify the amount of DP₁, DP₂, DP₃, DP₄ and above in the syrups. The chromatograms of GS and HMCS are depicted in Figure 4 and




Figure 4: Overlay chromatograms of 10 µL injections of 10 ppm DP₁—DP₃ mix (black), and 50 ppm GS sample (red).




Figure 5: Overlay chromatograms of 10 µL injections of 10 ppm DP₁—DP₃ mix (black), and 50 ppm HMCS sample (red).

Figure 5.

The amounts of DP₁—DP₃ were quantified using the calibration curve shown in figure 2, while DP₄ and above were quantified using the equation below. The average amounts based on triplicate measurements are shown in Table 5.

$$DP_4 \text{ and above} = TC - (DP_1 + DP_2 + DP_3)$$

TC = Total carbohydrate content on the label.

Table 5

| Average calculated amount of each DPs in the stock solutions of the sample, n = 3 | | |
|---|---------|-----------|
| Compound | GS (mg) | HMCS (mg) |
| DP ₁ | 4.29 | 8.20 |
| DP ₂ | 89.67 | 23.54 |
| DP ₃ | 57.08 | 41.50 |
| DP ₄ and higher | 16.21 | 98.36 |

The chromatogram of GS in figure 3 along with table 5 shows that GS primarily consists of DP₁—DP₃, with DP₂ being the most abundant component in the sample. In HMCS the DP₁ content is almost twice higher as in the GS sample. The DP₂ and DP₃ content in HMCS is significantly smaller compared to GS. DP₄ and higher DPs are the dominant components in the HMCS sample. The amount of DP₄ and higher DPs in HMCS is a factor of about 6× larger compared to the GS sample.



References

1. US FDA Code of Federal Regulations Title 21, 21CFR184.1444, www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfcfr/cfrsearch.cfm?fr=184.1444
2. Takeiti, C., Kieckbusch, T., Collares-Queiroz, F., Morphological and physicochemical characterization of commercial maltodextrins with different degrees of dextrose-equivalent, *Int. J. Food Propert.*, 2010, 13, 411-425.
3. Hofman, D. L., van Buul, V. J., Brouns, F. J. P. H., Nutrition, health, and regulatory aspects of digestible maltodextrins, *Crit. Rev. Food Sci. Nutr.*, 2016, 56(12), 2091-2100.
4. Børsheim, E., et al., Effect of carbohydrate intake on net muscle protein synthesis during recovery from resistance exercise, *J. Appl. Physiol.*, 2004, 96, 674-678.
5. Hansen, E. A., et al., Improved marathon performance by in-race nutritional strategy intervention, *Int. J. Sport Nutr. Exerc. metab.*, 2014, 24, 645-655.
6. BeMiller, J. N., Starches: Conversions, modifications, and uses, Chapter 7, *Carbohydrate Chemistry for Food Scientists* 3rd ed. Elsevier, 2019.
7. Dziejdzic, S. Z., Kearslye, M.W., *Handbook of starch hydrolysis products and their derivatives*, Blackie Academic & Professional, 1995, 230.
8. Rong, Y., Sillick, M., Gregson, C.M., Determination of dextrose equivalent value and number average molecular weight of maltodextrin by osmometry, *J. Food. Sci.*, 2009, 74(1), 33-40.
9. Moreno, F.J., et al., Determination of maltodextrins in enteral formulations by three different chromatographic methods, *Chromatographia*, 1999, 50, 705-710.
10. Moraes, F. S., et al., Development of HPLC-ELSD method for determination of maltodextrin in raw milk, *Food Chem.*, 2017, 217, 346-351.
11. Lee, Y.C., Carbohydrate analyses with high-performance anion-exchange chromatography, *J. Chrom. A.*, 1996, 720, 137-149.
12. ThermoFisher Scientific AppsLab Library, HPAE-PAD profiling of maltodextrins using dual eluent generation cartridges, <https://appslab.thermofisher.com/App/4300/hpaepad-profiling-maltodextrins>
13. Mechelke, M., et al., HPAEC-Pad for oligosaccharide analysis—novel insights into analyte sensitivity and response stability, *Anal. Bioanal. Chem.*, 2017, 409, 7169-7181.
14. Louw, H., Brouwer, H.J., Reinhoud, N., Electrochemical flowcell, US patent 9310330, 2016.
15. Rocklin, R.D., Clarke, A.P., Weitzhandler, Improved longterm reproducibility for pulsed amperometric detection of carbohydrates via a new quadruple-potential waveform, *Anal. Chem.*, 1998, 70, 1496-1501.

Conclusion

The ALEXYS™ Carbohydrate Analyzer in combination with SweetSep™ AEX200 anion-exchange column demonstrates a high-resolution separation of maltodextrins, offering insights into its composition up to DP₄₀. Demonstrated by the excellent linearity, repeatability, and sensitivity, the method provides a valuable tool for quantification of components in maltodextrins. The maltodextrin profile and the amount of DP₁, DP₂, DP₃, DP₄ and above in two syrup samples were assessed. In summary, the presented HPAEC-PAD method is a suitable method for the analysis of maltodextrins in food products.



Ordering information

| Detector only | |
|-----------------------------|--|
| 176.0035B | DECADE Elite SCC electrochemical detector |
| 116.4321 | SenCell 2 mm Au HyREF |
| Recommended ALEXYS analyzer | |
| 180.0057W | ALEXYS Carbohydrate Analyzer—gradient (quaternary LPG) |
| 116.4321 | SenCell 2 mm Au HyREF |
| 186.ATC00 | CT2.1 Column Thermostat |
| Column | |
| 260.0010 | SweetSep™ AEX200, 4 × 200 mm column, 5 µm |
| 260.0030 | SweetSep™ BIT, 4 × 50 mm borate ion trap |
| 260.0100 [#] | Pre-column filter PEEK, 0.5 µm |
| Software* | |
| 195.0035 | Clarity CDS single instr. incl. LC, AS module |

#) In case samples might contain particulate matter, it is advised to use a pre-column filter.

*) The ALEXYS Carbohydrate Analyzer can also be controlled under Thermo Fisher Scientific Chromeleon™ CDS. For the DECADE Elite electrochemical detector only, control drivers are also available for Waters Empower™, Agilent OpenLab CDS, and Agilent OpenLab CDS Chemstation Edition. Please Contact Antec for more details.

For research purpose only. The information shown in this communication is solely to demonstrate the applicability of the ALEXYS system and DECADE Elite detector. The actual performance may be affected by factors beyond Antec's control and may be adjusted accordingly. Specifications mentioned in this application note are subject to change without further notice.

SweetSep, DECADE Elite, ALEXYS, SenCell, FlexCell and HyREF are trademarks of Antec Scientific. Clarity™ and DataApex™ are trademarks of DataApex Ltd. Chromeleon™ is a trademark of Thermo Fisher Scientific, Empower™ is a trademark of Waters corporation, OpenLAB™ and Chemstation are trademarks of Agilent Technologies, Inc. All other trademarks are the property of their respective owners.

Antec Scientific (USA)
info@AntecScientific.com
www.AntecScientific.com
T 888 572 0012

Antec Scientific (worldwide)
info@AntecScientific.com
www.AntecScientific.com
T +31 172 268888

